

An aza-cyclophane stacked in racemic columnar assemblies: Whole molecular disorder in a 2D solid solution


Matthias Zeller,^{*#} Marlon R Lutz Jr.,^{\$} Daniel P. Becker^{*\$}

[#] Youngstown State University, Department of Chemistry, 1 University Plaza, Youngstown Ohio 44555-3663, USA

^{\$} Department of Chemistry, Loyola University, 6525 North Sheridan Road, Chicago, IL 60626

¹H-NMR of N-acetyl derivative 2 in d₆-DMSO with integration

ML6-070-4
Pulse Sequence: s2pul
Solvent: DMSO
Ambient temperature
File: ML6-70-4-1HMR-d6DMSO-Integrated
GEMINI-300 "gort.dpt-svr.luc.edu"
PULSE SEQUENCE
Relax, delay 1.000 sec
Pulse 31.5 degrees
Acq. time 1.330 sec
Width 4500.5 Hz
64 repetitions
OBSERVE H1, 300.0764735 MHz
DATA PROCESSING
FT size 32768
Total time 3 min, 19 sec



ML6-070-4

Pulse Sequence: s2pul

Solvent: DMSO

Ambient Temperature

File: ML6-70-4-1HMR-d6DMSO-Integrated

QEMINI-300 "gort.dpt-svr.luc.edu"

PULSE SEQUENCE

Relax. delay 1.000 sec

Pulse 31.5 degrees

Acq. time 1.958 sec


Width 4500.5 Hz

64 repetitions

OBSERVE H1, 300.0764795 MHz


DATA PROCESSING

FT size 32768



ML6-070-4

Pulse Sequence: s2pul
Solvent: DMSO
Ambient temperature
File: ML6-70-4-1HMR-d6DMSO-Integrated
GEMINI-300 "gort.dpt-svr.luc.edu"
PULSE SEQUENCE
Relax. delay 1.000 sec
Pulse 31.5 degrees
Acq. time 1.998 sec
Width 4500.5 Hz
64 repetitions
OBSERVE H1, 300.0764795 MHz
DATA PROCESSING
FT size 32768
Total time 3 min, 19 sec



ML6-070-4

Pulse Sequence: s2pu1

Solvent: DMSO

Ambient temperature

File: ML6-70-4-HNMR-460DMSO-Integrated

GEMINI-300 "gort.dpt-svr.luc.edu"

PULSE SEQUENCE

Relax. delay 1.000 sec

Pulse 31.5 degrees

Acq. time 1.998 sec

Width 4500.5 Hz


64 repetitions

OBSERVE H1, 300.0764735 MHz

DATA PROCESSING


FT size 32768

Total time 3 min, 19 sec




Sample: Becker

TGA File: C:\ta\Data\TGA\MRL\Becker




Sample: Becker

TGA File: C:\ta\Data\TGA\MRL\Becker



Curve 1: Becker1

DSC File: C:\ta\Data\DSC\Becker1



TGA

TGA data were collected on a 2050 Thermogravimetric Analyzer. Data were collected in open aluminum pans. Data were collected after equilibration at 28°C, followed by a 5°C/min ramp up to 200 °C.

The TGA shows loss of one chloroform molecule.

The MW of CTV-Amide N-Acetate CHCl₃ solvate = 640.936 g/mol

The MW of CTV-Amide N-Acetate = 521.558g/mol

The MW of CHCl₃ = 119.378 g/mol

Taking the difference between the CHCl₃ solvate and the CHCl₃ free CTV-Amide N-Acetate (640.936 g/mol - 521.558g/mol) gives 119.378 g/mol as expected.

Determining the weight loss: (119.378 g/mol) / (640.936 g/mol) x 100% = 18.6% weight loss.

Therefore, theoretically one should observe an 18.6% weight loss if CHCl₃ is the solvate. Furthermore, the corresponding weight percent of 81.4% (100 - 18.6%) reflects the solvate free CTV-Amide N-acetate.

The TGA confirms the theoretical weight loss percentage (18.6%) and the remaining weight (81.4%) calculated confirming that chloroform is lost and that the ratio of chloroform to substrate is 1:1.

Also, the TGA data relates well with the classical open-capillary melting point determination. That is if you look at the bell shaped curve, this tells us when CHCl₃ starts and finishes evolving. At about 115 deg C, one can see that sintering occurred. At 132 deg C where the TGA curve intersects the bell shaped curve, this indicates the phase change and noticeable CHCl₃ evolution. At 140 deg C, over half of CHCl₃ has evolved, and this is the point where the material resolidifies.

MP determination observed in an open capillary using an Electrothermal Mel-Temp:

116-118 C..... sintering

132-135 C.....a phase change occurred (melting) followed by refluxing (must be CHCl₃ boiling off)

140 C.....resolidification

175-178 C.....a second melting point

The TGA analysis also confirms that CHCl₃ must be boiling off at ~ 132 C. Only one CHCl₃ molecule is lost.

The TGA and X-ray results complement the H-NMR results in that ¹H-NMR (d6-DMSO) shows a 1:1 ratio of CHCl₃ to molecule **2**.

DSC

DSC data were collected on a 2910 Differential Scanning Calorimeter. Data were collected in sealed aluminum pans. The first measurement (Becker1) was conducted after equilibration at 28°C, followed by a 5°C/min ramp up to 200 °C, and recooling to 80 °C. The second measurement (Becker1.000) was conducted after equilibration at 28°C, followed by a 5°C/min ramp up to 110 °C, followed by further heating at 0.5°C/min and recooling to 80 °C, also at 0.5°C/min.

The 2nd melting point 175-178 C determined via capillary melting point is slightly different than the DSC results. The DSC data shows a melting point at about 187 C. The only two reasonable explanations of the two different results are:

- 1) the rate of heating
- 2) Closed system vs open system (Sealed pan vs open capillary). This would most likely give different results, because with the sealed pan, CHCl_3 is not lost, and with the capillary MP, CHCl_3 is lost into the atmosphere.