



Figure S4 Fluorescence microscopy set-up in the A1 station at CHESS utilizing a 5 mW 405 nm CW laser and an FGL435 Schott glass filter. The laser beam is aligned to the center of rotation of the spindle via a piece of paper mounted onto an ALS type base.