Qualitative Determination of Known Thromboxane Receptors in Human Plasma by LC-MS (Quadrupole Time of Flight)

KOTTE SUBHASH CHANDRA BOSE^{1,*}, P.K. DUBEY² and P.M. MURALI¹

¹Evolva Biotech Private Limited, TICEL Bio Park Limited, Taramani, Chennai-600 113, India

²Department of Chemistry, College of Engineering, Jawaharlal Nehru Technological University-Hyderabad, Kukatpally, Hyderabad-500 085, India

*Corresponding author: Tel/Fax: +91 44 42971050/60; E-mail: subhashcbk@gmail.com

(Received: 13 February 2013;

Accepted: 5 August 2013)

AJC-13891

A simple, rapid and sensitive reverse-phase LC-MS method has been developed for the simultaneous quantification of Dazmegrel, GR 32191, SQ-29548 and U-46619 in human plasma. Analytes were extracted from human plasma by liquid-liquid extraction technique using acetonitrile. Separation was achieved on a reverse phase Shodex C18, 3 μ m, 4.6 $\times$ 100 mm column at ambient temperature under isocratic conditions at a flow rate of 1 mL/min, run time of 10 min with acetonitrile and 0.1 % formic acid in water (70:30 v/v). Detection was performed using dual electrospray ionization (ESI) by using a fast polarity switch mode for fragment ions and adduct as for Dazmegrel, GR32191, SQ-29548 and U-46619 in positive and negative modes. The results indicate that this proposed method is convenient and reliable for clinical studies application.

Key Words: Dazmegrel, GR32191, LCMS, SQ-29548, Thromboxane receptors, U-46619.

INTRODUCTION

Thromboxane is a potent platelet aggregating agent which also contracts both vascular and non-vascular smooth muscle. Selective agents have been identified which inhibit these actions of thromboxane in a manner consistent with competitive receptor antagonism. Some of these antagonists have potential utility in the treatment of thromboxane-dependent pathologies, particularly within the cardiovascular/renal systems.

Classically, receptor types are defined by use of selective antagonists and possible receptor differences derived from differential potency of these antagonists when measured in different tissues. Dazmegrel¹⁻³ [3-(1*H*-imidazole-1-yl-methyl)-2-methyl-1*H*-indole-1-propanoic acid] is persistence as enzyme inhibitors; platelet aggregation inhibitors and Vasodilator Agents. GR32191⁴⁻¹⁰ (vapiprost) 7-(5-(((1,1'-biphenyl)-4-yl)-methoxy)-3-hydroxy-2-(1-piperidinyl)cyclopentyl)-4-heptanoic acid is a potent and selective thromboxane receptor antagonist. The thromboxane A₂/PG endoperoxide receptor inhibitor. Its mechanism is Fibrinolytic Agents; platelet aggregation inhibitors and prostaglandin antagonists. SQ 29548¹¹ 7-(3-(((phenylamino)carbonyl)hydrazine)methyl)-7-oxabicyclo-(2.2.1)hept-2-yl)-5-heptenoic acid is a Thromboxane/antagonists & inhibitors. U-46619¹²⁻¹⁶ (Z)-7-((1*R*,4*S*,5*S*,6*R*)-6-((*E*)-3-hydroxyoct-1-enyl)-2-oxabicyclo[2.2.1]heptan-5-yl)hept-5-

enoic acid has been postulated to inhibit platelet aggregation by competitively interfering with the thromboxane A₂ (TxA₂) platelet receptor, which was found to be able to displace the platelet binding and a labeled TxA₂ antagonist. A stable prostaglandin endoperoxide analog which serves as a thromboxane mimetic. Its actions include mimicking the hydro-osmotic effect of VASOPRESSIN and activation of TYPE C PHOSPHOLIPASES.

Literature survey revealed that no method has been reported till date for determination of Dazmegrel, GR 32191, SQ-29548 and U-46619 in human plasma. As per ICH guidelines here we are reporting the availability of structurally different thromboxane receptor antagonists has allowed analysis of thromboxane receptors characteristics in human plasma by using Accurate Mass Q-TOF LC/MS. The proposed LCMS method was validated by assessing its specificity, linearity, accuracy, precision, limits of detection and quantification, system suitability parameters, ruggedness and robustness. This article reports for the first time the results of a study carried out to develop a straightforward HPLC PDA-ESI-MS method, which employs liquid-liquid extraction technique for sample preparation and simultaneous identification and quantification of Dazmegrel, GR 32191, SQ-29548 and U-46619 in human plasma, thus the proposed method can be used for routine analysis in clinical trials.

Structure of thromboxane receptors (1) Dazmegrel, (2) GR32191, (3) SQ-29548 and (4) U-46619

EXPERIMENTAL

Pure standards of SQ-29548, U-46619, GR32191 and Dazmegrel were obtained from the TOCRICS and ALEXIS. All the chemicals used were of analytical reagent grade and the solvents were of ACS in LCMS grade, HPLC Ultra Gradient Solvent Acetonitrile BAKER ANALYZEDR® was purchased from J.T. Baker, Avantor Performance Materials (USA), ultrapure water was obtained using an Elix Advantage 5 system coupled to a Milli-Q Biocel (Billerica, MA, USA). All solutions were degassed by ultrasonic cleaner using KUDOS (Shanghai, P.R.China) and filtered through Millex-FG Filter, 0.20 μ m, Fluoropore™ (PTFE) 13 mm, non-sterile membrane sample filter before injecting into system. Drug-free human plasma was obtained from the Holy mother diagnostics and Blood Bank (Madipakkam, Chennai, India).

LC-MS conditions and accurate-mass spectrometry quadrupole time of flight: The LC method was carried out using an Agilent 1200 series HPLC system (Agilent Technologies, USA), equipped with a binary pump, a degasser, PDA detector, an auto-sampler, a column oven. The separation was carried out on a reverse phase Shodex C18 - 4C, 3 μ m, 4.6 $\times$ 100 mm at ambient temperature. A security guard holder was used to protect the analytical column. Isocratic elution was employed using 0.1 % formic acid in water (solvent A) and Acetonitrile (solvent B) (70:30 (v/v) and eluted by the following program at the flow 1 mL/min with run time of 10 min.

Agilent 6520 Quadrupole time-of-flight (Q-TOF) mass spectrometer coupled to an Agilent 1200 series HPLC system. The Agilent's Q-TOF mass spectrometer is equipped with dual electrospray ionization (ESI) ion source. Precisely mass spectra were acquired by using fast polar switching mode with scan range from m/z 100 to 600 Da. The conditions of dual ESI source were as followed: drying gas (N_2) flow rate, 30 L/min;

temperature, 350 $^{\circ}$ C; pressure of nebulizer, 50 psi; capillary voltage, 3500 V, drying gas 10 L/min and fragmentor voltage, 175 V and MS parameters are showed in Table-1. Data were acquired and analyzed by Agilent Mass Hunter Workstation Software version B.02.01 (B2116.20) and Quantitative Analysis Version B.01.04 (B 1.4.126.0) (Agilent Technologies, USA). The output signal is monitored and processed using mass hunter software on Intel® Core (TM) 2 Duo computer (HP xw 4600 Workstation). This instrument was used to confirm the identification of chromatographic peaks of interest.

TABLE-1
MS PARAMETERS

Capillary voltage (V cad)	3500 V
Fragmentor	175 V
Gas flow	10 L/min
Gas temperature	350 $^{\circ}$ C
Ionization mode	Dual ESI fast polarity
Nebulizer	50 psi
Octo dole RF peak	750 V
Scan rate	0.8 spectra/s
Skimmer	65 V

Preparation of standard stock solutions: Mixed standard stock solutions were prepared by accurately weighed the all standards and dissolved in suitable solvent (methanol). The working standard solution was prepared by diluting the mixed standard solution with the same to a series of proper concentrations. The standard stock and working solutions were all stored at 4 $^{\circ}$ C until use.

Preparation of plasma samples: A 50 μ L aliquot of the premix stock solution was added into 200 μ L of drug free human plasma and samples were mixed for 3 min by vortex, add 200 μ L of acetonitrile to crash proteins and centrifuged at 14000 rpm for 10 min. The organic layer was transferred to a

test tube and evaporated to dryness under a stream of nitrogen at 40 °C. The residue was reconstituted in 100 µL of mobile phase. After centrifugation at 14000 rpm for 2 min, 2 µL of the supernatant was subjected to analysis.

Calibration curves: The working standard solutions were brought to room temperature and an aliquot of 2 µL was injected into LC-MS for the construction of calibration curves. Eleven concentrations in triplicate were analyzed and the calibration curves were calculated by linear regression of the double logarithmic plots of the peak area *versus* the amount of thromboxane receptors injected.

Limits of detection and quantitation: The limits of detection (LODs) and quantification (LOQs) under the present chromatographic conditions were determined by diluting the standard solution when the signal-to-noise ratios (S/N) of analytes were almost 3 and 10, respectively. The S/N was calculated as the peak height divided by the background noise value. The background noise was measured from the background start to background end time.

Precision, repeatability and accuracy: Intra and inter-day variations were chosen to determine the precision of the developed method. For intra-day variability test, the working standard solutions (at low, medium and high levels of concentration) were analyzed in triplicate three times within one day, whereas for inter-day variability test, the working solutions were examined in triplicate for consecutive 3 days. Variations of the peak area were taken as the measures of precision and expressed as percentage relative standard deviations (R.S.D.).

For repeatability test, six independent analytical sample solutions from the same batch were used. R.S.D. (%) values of the obtained contents of each analyte were used to estimate repeatability.

Robustness: Robustness of the method was studied by deliberately changing the experimental conditions like flow rate, percentage of organic phase, pH of mobile phase and also by observing the freeze and thaw process of the sample solution at 0 ± 2 °C for 24 h.

RESULTS AND DISCUSSION

Method development by LC-MS-dual ESI: The novel and new chromatographic conditions for determination of thromboxane receptors by HPLC- ESI-MS were used and mass parameters were tuned to fast polarity switching mode for the analytes. Good response was found in positive ionization mode for Dazmegrel, GR32191 and SQ-29548 and negative ionization mode for Dazmegrel, GR32191, SQ-29548 and U-46619. For the purpose of correct identification, a LCMS analysis

was performed on sample solutions under the dual ESI-MS fast polarity conditions. Analytes, retention time, fragment ion and adduct ions are summarized in Table-2. In positive ion mode, the compounds of interest exhibited mainly protonated ions and adduct ions. Finally, the identified thromboxane receptors by comparing their retention times and MS data with those of reference compounds.

Chromatographic conditions, precisely the composition of the mobile phase, column type, flow rate and column oven temperature was optimized through numerous trials to obtain good resolution and amplified intensity of the signals of the analytes, as well as for the short run time. The presence of a small amount of formic acid in the mobile phase improved the detection of the analytes. It was found that a mixture of the isocratic mobile phase was 0.1 % formic acid in water and acetonitrile (30:70 v/v) could achieve this purpose and was finally adopted as the mobile phase. Shodex C18, 3 µm, 4.6 × 100 mm column gave good peak shape and response even at lowest concentration level for the analytes. The mobile phase was operated at a flow rate of 1 mL/min. The retention time of Dazmegrel, GR32191, SQ-29548 and for U-46619 (4.237, 6.635, 7.855 and 8.570 min, respectively) were low enough allowing a run time of 10 min.

Liquid-liquid extraction (LLE) technique was applied for the sample preparation in this work. Liquid-liquid extraction is helpful in producing a spectroscopically clean sample and avoiding the overview of non-volatile components onto the column and MS system and also reducing the experimental cost. System suitability parameters like coefficient factor, resolution, tailing factor, theoretical plates and symmetry are within the range and are described in Table-3. Among the different solvents checked alone and in combination for their suitability, acetonitrile and water was found to be optimal, which can produce a clean chromatogram for a blank sample and yields the highest recovery for the analytes from the plasma.

Linearity, limit of detection and limit of quantification: Linearity was demonstrated by using mass chromatographic data by extraction ion current chromatogram from 10 ppm to 125 ppm of the standard concentration in mixture form using eleven calibration levels (10, 20, 30, 40, 50, 60, 70, 80, 90, 100 and 125 ppm) for Dazmegrel, GR32191, SQ-29548 and U-46619 (Fig. 1). The method of linear regression was used for data evaluation. The peak areas of the standards in mixture form were plotted against the respective standard concentration. Linearity was described by the linearity equation, the correlation coefficient and excellent linearity obtained in above concentrations levels.

TABLE-2

COMPOUNDS SEPARATED (ESI-MS DETECTION), MOLECULAR FORMULA AND MOLECULAR WEIGHT FOR EACH ANALYTE

Peak No.	Compound	t _R (min)	Molecular mass	Fragment ion m/z	Ion	Adduct ion m/z	Ion
1	Dazmegrel	4.223	283.16246	282.15197	(M-H)-	565.31244	2(M-H)-
				284.17010	(M+H)+	567.33745	2(M+H)+
2	GR32191	6.651	477.34243	476.32834	(M-H)-	512.30818	(M-Cl)-
				478.52128	(M+H)+		
3	SQ-29548	7.533	387.25943	386.24580	(M-H)-	422.22686	(M-Cl)-
				388.26556	(M+H)+	410.25077	(M+Na)+
						426.22633	(M+K)+
4	U-46619	9.406	350.27918	349.27167	(M-H)-	385.25203	(M+Cl)-

Fig. 1. HPLC-ESI-MS for the samples Dazmegrel, GR32191, SQ-29548 and U-46619 with Extraction ion current chromatogram (EIC) acquired in negative polarity scan mode and its ion chromatograms extracted at m/z values corresponding polarity

Compound	Dazmegrel	GR 32191	SQ-29548	U-46619
RT	4.237	6.635	7.855	8.57
Area	1581977	999300	2057000	936532
Base Peak	565.3119	476.32867	386.24651	384.23146
k'	-0.6	-0.3	-0.2	-0.1
Resolution	51.1	14.5	2.2	2.6
Tailing factor	1	1.5	1.3	1.2
Plates	10437	25816	45679	59845
Plates/M	69580	172106.7	304526.7	398966.7
Symmetry	0.67	0.33	1	1
SNR	143.3	111.2	226.1	114.7

As shown in Table-4, acceptable results of the regression analysis, the correlation coefficients (r^2), LODs and LOQs were obtained for all the analytes: all calibration curves showed good linear regression within the test ranges; the LODs and LOQs of the Dazmegrel, GR32191, SQ-29548 and U-46619 were in the range of 7.322-24.405 $\mu\text{g/mL}$, 15.935-53.116 $\mu\text{g/mL}$, 12.016-40.053 $\mu\text{g/mL}$ and 19.248-64.161 $\mu\text{g/mL}$, respectively.

Precision and accuracy: Instrument precision was determined by performing injection repeatability test and the RSD value for Dazmegrel, GR32191, SQ-29548 and U-46619. The intra-day precision studies were carried out and RSD

values were found to be 0.15-0.81, 0.43-0.62, 0.4-0.99 and 0.61-1.2 %; the inter-day precision studies were carried out and RSD values were found to be 0.61-1.54, 0.62-1.00, 0.38-1.7 and 76-1.14 %, respectively (Table-5). The low RSD values indicate that the method is precise.

The accuracy of the method was determined by calculating recoveries of Dazmegrel, GR32191, SQ-29548 and U-46619 by method of standard addition. The recoveries found to be 94.42-101.08, 86.18-111.17, 90.47-109.44 and 103.93-109.45 %, respectively (Table-6). The high values indicate that the method is accurate.

Conclusion

The proposed study describes qualitative LCMS method for estimation of Dazmegrel, GR32191, SQ-29548 and U-46619 in drug free human plasma. The method was validated and found to be very simple, sensitive accurate and precise. The simple liquid-liquid extraction method gave consistent and reproducible recoveries for the analytes from plasma. The method provided good linearity. The method is sensitive enough for quantitative detection of the analytes. Thus the proposed method can be used for routine analysis in clinical trials.

ACKNOWLEDGEMENTS

The authors are grateful to the Evolva Biotech Pvt. Ltd., India for providing laboratory facility for this research work

Analyte	Calibration curve	Linearity r^2	Limit of detection ($\mu\text{g/mL}$)	Limit of quantification ($\mu\text{g/mL}$)
Dazmegrel	$Y = 17436.389154x + 139624.144424$	0.95915301	7.322	24.405
GR32191	$Y = 10992.180210x + 74859.317172$	0.97292508	14.935	53.116
SQ-29548	$Y = 27724.437682X - 26335.089848$	0.97829994	12.016	40.053
U-46619	$Y = 9498.000625x - 132406.920330$	0.95860995	19.248	64.161

TABLE-5
PRECISION AND ACCURACY DATA FOR THXA2 IN HUMAN PLASMA SAMPLES

Analyte	Concentration (ppm)	Intra-day precision and accuracy (n = 6)			Inter-day precision and accuracy (n = 6)		
		Concentration found	% RSD	Accuracy (%)	Concentration found	% RSD	Accuracy (%)
Dazmegrel	10	10.6798	0.49	106.8	10.7885	0.61	107.88
	20	21.5980	0.15	107.99	21.7469	1.54	108.73
	30	30.6078	0.65	102.03	29.0150	0.85	96.72
	40	39.4505	0.77	98.63	33.2873	0.77	83.22
	50	53.7489	0.71	107.5	54.5627	0.80	109.13
	60	62.6669	0.81	104.44	66.6823	1.14	111.14
GR32191	10	9.9077	0.46	89.08	10.0560	0.62	100.56
	20	22.1153	0.51	110.58	19.2492	1.00	96.25
	30	32.3964	0.62	107.99	29.7122	0.69	99.04
	40	38.0847	0.43	95.21	44.4189	0.97	111.05
	50	49.7097	0.51	99.42	49.2728	0.82	98.55
	60	60.5120	0.49	100.85	68.1045	0.74	113.51
SQ-29548	10	9.0090	0.99	99.08	10.5151	1.70	105.15
	20	18.9559	0.67	94.78	20.8421	0.40	104.21
	30	30.0451	0.64	100.15	30.9363	0.82	103.12
	40	36.1346	0.40	90.34	32.0568	0.38	80.14
	50	52.5314	0.69	105.06	54.2367	0.61	108.47
	60	63.6445	0.87	106.07	63.9918	0.98	106.65
U-46619	10	11.1123	0.61	111.12	11.9183	0.86	119.18
	20	17.0481	1.20	85.24	20.5693	0.75	102.85
	30	26.4533	0.89	88.18	24.6241	0.88	82.08
	40	34.4475	0.97	86.12	38.8392	0.76	97.10
	50	45.2124	0.89	90.42	43.5963	1.02	87.19
	60	55.7526	0.93	92.92	53.5268	1.14	89.21

TABLE-6
RECOVERY DATA OF STANDARD SOLUTIONS
ADDED TO THE SAMPLES ANALYZED USING
THE PROPOSED HPLC METHODS

Analyte	Added (µg/mL)	Measured (µg/mL)	Recovery % (mean ± SD)
Dazmegrel	100	100.2638	100.26 ± 9.85
	125	118.0194	94.42 ± 15.82
	250	252.7091	101.08 ± 10.00
GR32191	100	86.1809	86.18 ± 10.41
	125	125.7788	100.62 ± 20.82
	250	277.9137	111.17 ± 17.32
SQ-29548	100	105.1364	105.14 ± 11.55
	125	136.8009	109.44 ± 20.00
	250	226.1709	90.47 ± 15.28
U-46619	100	103.9274	103.93 ± 15.28
	125	136.8080	109.45 ± 20.01
	250	265.9903	106.40 ± 20.13

and Holy Mother Diagnostics and Blood Bank (Chennai, India) for providing human plasma.

REFERENCES

1. M.V. Rebec and V.A. Skrinska, *J. Chromatogr.*, **432**, 375 (1988).
2. G.T. Swayne, J. Maguire, J. Dolan, P. Raval, G. Dane, M. Greener and D.A. Owen, *Eur. J. Pharmacol.*, **152**, 311 (1988).
3. R.P. Gladue and M.F. Newborg, *Transplantation*, **52**, 837 (1991).
4. P. Lumley, B.P. White and P.P. Humphrey, *Br. J. Pharmacol.*, **97**, 783 (1989).
5. M.R. Foster, E.J. Hornby and L.E. Stratton, *Thromb. Res.*, **65**, 769 (1992).
6. M. Thomas, R.J. Keery, M.K. Charter, N.L. Scully, J.E. Chilton and P. Lumley, *Br. J. Clin. Pharmacol.*, **32**, 181 (1991).
7. R.L. Lorenz, S. Fischer, W. Wober, H.A. Wagner and P.C. Weber, *Biochem. Pharmacol.*, **35**, 761 (1986).
8. J.M. Ritter, N. Benjamin, H.S. Doktor, S.E. Barrow, T.G. Mant, S. Schey and P. Stewart-Long, *Br. J. Clin. Pharmacol.*, **29**, 431 (1990).
9. T. Uematsu, S. Nagashima, A. Mizuno, K. Hirano and M. Nakashima, *J. Clin. Pharmacol.*, **31**, 815 (1991).
10. R.C. Beasley, R.L. Featherstone, M.K. Church, P. Rafferty, J.G. Varley, A. Harris, C. Robinson and S.T. Holgate, *J. Appl. Physiol.*, **66**, 1685 (1989).
11. D. Zhu, M. Bousamra, D.C. Zeldin, J.R. Falck, M. Townsley, D.R. Harder, R.J. Roman and E.R. Jacobs, *Am. J. Physiol. Lung Cell Mol. Physiol.*, **278**, L335 (2000).
12. D.W. Phillipson, A.A. Tymiak, J.G. Tuttle, K.S. Hartl, T.W. Harper, M.S. Bolgar, G.T. Allen and M.L. Ogletree, *J. Lipid Mediat.*, **7**, 155 (1993).
13. C.H. Yeh, H.C. Peng, J.B. Yih and T.F. Huang, *Biochim. Biophys. Acta*, **1425**, 493 (1998).
14. P.A. Modesti, A. Colella, R. Abbate, G. Gensini and G.N. Serneri, *Eur. J. Pharmacol.*, **169**, 85 (1989).
15. S.J. MacLennan and G.R. Martin, *Br. J. Pharmacol.*, **107**, 418 (1992).
16. E.J. Kattelman, D.L. Venton and G.C. Le Breton, *Thromb. Res.*, **41**, 471 (1986).