Development of Electrochemical Immunosensor for the Detection of Human interleukin-37 for Potential Diabetes Diagnosis

Ge Zhang^{1,2}, Wei Huang², Haojun An², Chunjun Li³ and Demin Yu^{3,*}

¹ Tianjin Medical University, Tianjin, 300070, P.R. China

² Tangshan Worker Hospital, Tangshan, 063000, P.R. China

³ Metabolic Disease Hospital of Tianjin Medical University, Tianjin, 300070, P.R. China

*E-mail: deming_tianjing@qq.com

doi: 10.20964/2017.03.04

Received: 25 November 2016 / Accepted: 28 December 2016 / Published: 12 February 2017

Interleukin-37 (IL-37) has been sensitively detected by the electrochemical immunosensors owing to the dual amplification of AuNP-PDA (gold nanoparticles and Polydopamine). PDA was employed as an immobilization biomolecules not only to construct a sensor platform but also to label signal. The amplification of the sensitivity can be achieved by the employment of particular AuNP-PDA platform and AuNP-PDA@graphene modified by HRP-Ab2 (horseradish peroxidase and antibody). The response of IL-37 was evaluated by the determination of amperometry, and the linear range is starting from 4 pg/mL to 800 pg/mL and the detection limit was very low (0.8 pg/mL). The performance of immunosensors for detecting IL-37 in human serum was compared with assays of standard ELISA. Furthermore, the diagnosis of clinic diabetes would be potentially achieved by this well-defined immunosensor.

Keywords: Immunosensor; Human interleukin-37; Diabetes diagnosis; Polydopamine; AuNP

[FULL TEXT](#)

© 2017 The Authors. Published by ESG (www.electrochemsci.org). This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution license (<http://creativecommons.org/licenses/by/4.0/>).