iJRASET

International Journal For Research in
Applied Science and Engineering Technology

INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Volume: 9   Issue: VIII   Month of publication: August 2021

DOI: <https://doi.org/10.22214/ijraset.2021.37644>

www.ijraset.com

Call: 08813907089

E-mail ID: ijraset@gmail.com

Design and Fabrication of Wheelchair into Foldable Bed

Vinay D. R¹, Asutosh Kumar Sah², Washiuddin Khan³, Manish Kumar Chaudhary⁴, Bishal Pulami Magar⁵

¹Assistant Professor, Department of Mechanical Engineering, New Horizon College of Engineering, Bangalore-560103, Karnataka, India

^{2, 3, 4, 5} Students, Department of Mechanical Engineering, New Horizon College of Engineering, Bangalore-560103, Karnataka, India

Abstract: In this paper lead screw mechanism is used to convert a wheelchair into bed and vice versa. When lead screw mechanism is operated, the foot rest which is downward goes till the level of seat of wheelchair and simultaneously the back support goes down to the same level making, its stretcher and vice versa, when it is converted into wheelchair from stretcher. Below the seat, a defecation system has been provided which is detachable and is operated by sliding mechanism. It is designed and fabricated in such a way that it can be used in hospitals as well as for personal use at home. Below the seat, a defecation system has been provided which is detachable and is operated by sliding mechanism. Also, obstacle avoiding sensor i.e., ultrasonic sensor is used to know if any obstacle will come in front of wheelchair, siren bell will ring.

Keywords: Fabrication, Hospital, Mobility Aid, Lead Screw, Design, Wiper Motor, Wheelchair, Battery.

I. INTRODUCTION

Design an automatic wheelchair is the important things present day. In this generation, for the disabled people where the care requires a lot of money, labour and time. This report is the example of a design of an automated wheelchair into a foldable bed that will be able to perform all the functions in these modern wheelchairs (Wheelchair with the adjustable portion of the backrest and also footrest). Physically disabled people are affected millions of families in the world. In the present century, million people are suffering from physical challenged.

The disabilities of people are often in the empowered and enable them to measure a traditional life a normal and independent life with the assistance of wheelchair. Latest and modified wheelchairs can give the comfort and ease to disable people instead of the old peoples. It will help the disabled people in various design changes are to be done. It is expected that these new automatic wheelchair into foldable bed, sections (legs positions will be adjusted). It is expected that the new type of automatic wheelchair may enable people's better in the medical care of the patient, and would greatly decrease time and labours to the old-age home staff and also help for physical disabled people. It gives easy and comfort to the patients and the medically staffs.

II. LITERATURE SURVEY

The paper titled "Design and fabrication of the multiutility wheel-chair". In this paper, wheel-chair that can prevail over this conventional wheel-chair. By adopting different research, papers aided to spark the varied problems with objects regarding the solutions, precautions and general advancements and propelled of different materials for the development of wheel chair and ideas of manufacture regarding the advancement of power chairs. Generally, the project focuses on cost effectiveness and easily acceptance.

This Designed of Wheel-Chair Stretcher enables the easy to transferring and handling of patients in hospitals without producing any damage to patient's body externally and internally. Thus, the time and effort required for moving the patient may be reduced. Modelling software such as cad has helped in visualizing the product.

III. METHODOLOGY

The model works on lead screw mechanism. A lead screw or power screw or translation screw is a screw used as a linkage in a machine to translate turning motion into linear motion. Because of the large area of sliding contact between their male and female members, screw threads have larger frictional energy losses compared to other linkages. They are not typically used to carry high power, but more for intermittent use in low power actuator and positioned mechanisms.

Fig.1: flowchart of methodolog

IV. COMPONENT USED IN FABRICATION

Actually, wiper motors are generally design for the 2-speed operation process. The wiper motor mainly consists of three brushes namely; common, low and high-Speed motors. These two types of the brushes may supply for vary mode of operation process. 12v high torque low rpm electric motor.

Structure: synchronous motor

Shape: tabular

Phase: single-phase

Speed: High speed

Function: control

Power output: 25-30 W

Current: 10-13 Amps

Torque: 18.5 N-m

Load speed: 85 ± 15 rpm

Current (No load): < 5 A

V. COMPONENTS DESIGNED IN CATIA V5

Fig. 2 Wiper Motor

Fig. 3 Rear Wheel

Fig.4 cater

Fig.5 Caster Frame

VI. ASSEMBLED DESIGN AND IT'S DRAFTING

Fig.6 3D Design

Fig.7 Isometric View

TABLE I
Dimensions of Wheel Chair

Dimensions	Centimetre (cm)
Overall length	198 cm
Overall height	135 cm
Height from bottom to seat	76 cm
Overall width	69 cm

VII. DESIGN CALCULATION

A. Calculation of Input Power

We will use International Standard of units. By the definition of Ohm's Law. Ohms' law states that the generated current passing through the conductor is directly proportional to applied voltage, which can be expressed in the form of current, and voltage

Current (I) = Voltage (V) /Resistance(R)

Current (I) = 12 / Resistance(R)

Resistance(R) = 12/1.2

Resistance(R) = 10 Ohms

The electrical power can be calculated by using these formulae:

$P_{in} = \text{current}(I) * \text{Voltage}(V)$

$P_{in} = 12 * 12$

$P_{in} = 144 \text{ W}$

B. Calculation Of Output Power And Angular Speed

Input power (Pin) is measured in Watts

Current (I) is measured in Amperes

Applied voltage (V) is measured in Volts

Mechanical Power Output for the motor can be calculated by using the formulae:

$$P_{out} = \text{Torque}(\tau) * \text{Angular speed}(\omega)$$

$$P_{out} = 18.5 * \text{Angular speed}(\omega)$$

$$\text{Angular speed}(\omega) = ?$$

Where;

Pout – power output in Watts

Torque (τ) in Newton meters

Angular speed (ω) in radian per second

$$\text{Angular speed}(\omega) = \text{rotational speed} * 2\pi / 60$$

$$\text{Angular speed}(\omega) = 63 * 2\pi / 60$$

$$\text{Angular speed}(\omega) = 6.5973 \text{ rad/sec}$$

$$P_{out} = 18.5 * 6.5973$$

$$P_{out} = 122 \text{ W}$$

Where;

π -Mathematically constant pi (3.14)

60 seconds in a minute

By using mathematical formula to get efficiency

$$\text{Efficiency (E)} = \text{Power out (Pout)} / \text{Power in (pin)}$$

After substitution we get

$$E = P_{out} / P_{in}$$

$$E = 122 / 144$$

$$E = 0.847 * 100$$

$$\text{Efficiency} = 84\%$$

C. Load Calculations

1) Front Caster

Weight of the wheel chair = 24.45 Kg

Weight of the body = 80 Kg

Vertical force = (239.61 + 784) Kg

Load of each caster= (1023.61)/2 Kg

Force on each caster= (909.5)/2 Kg

2) Back Rest

Human back weight = 310.91 N

Incline = 25degree

Back rest weight = 58.8 N

3) Leg Rest

Inclination = 55 degree

Weight of the leg = 4 kg

Entire force = 39.2+39.2

$$= 78.4 \text{ Newton}$$

$$\text{Actual force} = 78.4 * \sin(50) = 64.22 \text{ Newton}$$

D. Calculation of Lead screw

$$\begin{aligned}\text{Linear velocity of the lead screw} &= \text{Number of rotations} \times \text{pitch diameter} \\ &= 50 \times 10 \text{ millimetres (mm)/minutes (min)} \\ &= 500 \text{ millimetres (mm)/minutes (min)} \\ &= 8.33 \text{ millimetres (mm)/seconds (sec)}\end{aligned}$$

$$\begin{aligned}\text{Angular velocity of lead screw} &= (2 \times \pi \times N) / 60 \\ &= (2 \times \pi \times 50) / 60 \\ &= 5.326 \text{ radians/seconds}\end{aligned}$$

$$\begin{aligned}\text{Pitch diameter (D)} &= (d_1 + d_2) / 2 \\ &= (14 + 18) / 2 \\ &= 16 \text{ mm}\end{aligned}$$

$$\begin{aligned}\text{Helix angle } (\alpha) &= \tan^{-1} \left(\frac{1}{\pi \times d} \right) \\ &= \tan^{-1} \left(\frac{1}{\pi \times 16} \right) \\ &= 11.25 \text{ degree}\end{aligned}$$

$$\begin{aligned}\text{To lift the load, torque (T)} &= W \times (D/2) \times \tan(\beta + \alpha) \\ &= 784.8 \times 2 \times \tan(12.95 + 11.23) \\ &= 268.42 \text{ N-mm}\end{aligned}$$

$$\begin{aligned}\text{Power of the lead screw} &= T \times [2 \times \pi \times N] / 60 \\ &= 0.268 \times [2 \times \pi \times 50 / 60] \\ &= 1.4 \text{ Watts}\end{aligned}$$

$$\begin{aligned}\text{Torsional moments (MT)} &= W \times (d_2/2) \times \tan(\beta + \alpha) \\ &= 80 \times 9.81 \times 9 \times \tan(11.25 + 12.95) \\ &= 3174.32 \text{ N-mm}\end{aligned}$$

$$\begin{aligned}\text{Torsional shear stress } (\tau) &= (16 \times 3.174) / (\pi \times 143) \\ &= 5.89 \times 10.3 \text{ N/mm}^2\end{aligned}$$

$$\text{Principle shear stress } (\tau_{\max}) = 2.549 \text{ N/mm}^2$$

$$\text{Principle normal stress } (\sigma_{\max}) = 5.098 \text{ N/mm}^2$$

$$\begin{aligned}\text{Transverse shear stress } (\tau_s) &= W / (\pi \times d_1 \times t \times i) \\ &= (80 \times 9.81) / (\pi \times 14 \times 4 \times 43) \\ &= 0.103 \text{ N/mm}^2\end{aligned}$$

$$\begin{aligned}\text{Efficiency (E)} &= \tan(\alpha) / \tan(\phi + \alpha) \\ &= \tan(11.25) / \tan(12.95 + 11.25) \\ &= 0.44 \\ &= 44\%\end{aligned}$$

If efficiency of screw < 50 %

Then it will be self -locked to the system.

1.5. SPEED CALCULATION OF WHEELCHAIR

$$\text{Radius of rear wheel (R)} = 355.6 \text{ mm}$$

$$\begin{aligned}\text{Perimeter (P)} &= 2 \times \pi \times R \\ &= 2234.30 \text{ mm}\end{aligned}$$

$$N_1 = 85 \text{ rpm}$$

$$\begin{aligned}V_1 &= P \times N_1 \\ &= (2234.30 \times 85) / (60 \times 1000) \\ &= 3.16 \text{ mm/sec}\end{aligned}$$

VIII. TESTING

After the fabrication, we have performed the various test like speed of the wheelchair without load and the speed of the wheelchair with the load (person) of different weight and found that the speed was without load and with load. Also, the transformation of wheelchair into bed and vice versa was successful without load and with load. We have successfully installed the Ultra-Sonic sensor which detect obstacle on the path and alert. The wheelchair was comfortable and easy to handle. The battery backup was also good.

Fig.8 Front View

Fig. 9 Side View

Fig.10 Wheelchair into Bed

IX.CONCLUSION

This wheel chair has been designed in order to reduce the pain for patients and pregnant woman too. This wheel chair can be easily afforded new design. This can be generally helpful in hospital, health-post, old age home and also for disabled societies. Mostly, this wheel chair will be most helpful to prenatal and postnatal too. This wheel chair will be more advanced technologies. we have come to that the analysis of patients in the hospitals, health-post, old-age home and the analysis of medical market. To overcome these problems regarding the patients, we have designed and fabricated a wheel chair into foldable bed. Make sure this wheel chair into foldable bed may help the victims for relaxing and relief over the pain and problems.

REFERENCES

- [1] Sandip S. Bag, Prem D. Lohe, Harshal K. Hajare, A. N. Madne and Ajinkya S.Hande, "Design and fabrication of multiutility wheelchair", [2017].
- [2] SukantaRoga, Abhijeet Kumar, Animesh Singh, Bijesh Kumar. Aman, "Design and Fabrication of Wheelchair cum Stretcher with Multi Fold", [Jun 2017].
- [3] Sandeep Kumar, Gaurav Kaintura, Gurpreet Singh, "Fabrication of Collapsible Chair Cum Bed ModelAn Engineering", [Aug 2018]
- [4] Arun Kumar S, Boney Philip Parampil, Febin P Sahib Nazeer, Febin Stephen Joseph, "Design and Fabrication of Multipurpose Mobility Carriage for Crippled", [Jun 2018]
- [5] Vikas P. Mapari, Ankush S. Mashakhtri, Ajinkya S. Hande, "A Review on Design and Fabrication of Multiutility Wheel Chair". [Mar 2017]
- [6] F. Capezio, F. Mastrogiovanni, A. Scalmato, A. Sgorbissa, P. Vernazza, T. Vernazza, R. Zaccaria," Mobile Robots in Hospital Environments: an Installation Case Study". [2017]
- [7] Laurel D. Riek, Healthcare Robotics. Apr 2017
- [8] Dr. K. Lakshmi Narayanan, Dr. N. Muthu Kumaran, Dr.G.Rajakumar, : Design and Fabrication of Medicine Delivery Robots for Hospitals. [Oct 2019]
- [9] Richard C. Simpson Phd, "Smart Wheelchairs", Department Of Rehabilitation Science And Technology, University Of Pittsburgh,Pa. [2005]

10.22214/IJRASET

45.98

IMPACT FACTOR:
7.129

IMPACT FACTOR:
7.429

INTERNATIONAL JOURNAL FOR RESEARCH

IN APPLIED SCIENCE & ENGINEERING TECHNOLOGY

Call : 08813907089 (24*7 Support on Whatsapp)