AN ELECTROCHEMILUMINESCE MICRO-INSTRUMENT

Yun-Tai Hsueh and Rosemary L. Smith

Department of Engineering Electrical And Computer Science, University of California, Davis
Davis, CA 95616

M. Allen Northrup

Microtechnology Center, Lawrence Livermore National Laboratory
Livermore, CA

ABSTRACT

This paper describes an electrochemi-luminescence (ECL) instrument, which is micro-fabricated in Silicon and demonstrates its performance. The instrument contains an ECL excitation cell, fluid fill ports and a PIN silicon diode detector. The detection limit at present is tens of atto(10^{-18}) moles of $\text{Ru}^{+2}(\text{bpy})_3$.

INTRODUCTION

The ECL microinstrument was designed for DNA quantification after amplification by Polymerase Chain Reaction (PCR). The detection limit is desired to be as low as possible to minimize the amplification and total analysis time. To do so, a high gain, low noise detector such as a photomultiplication tube (PMT), cooled avalanche photodiode, etc., are normally employed. However, our aim is to make a portable instrument, preferably hand-held, therefore size and power requirements exclude those detectors. Here a simple, small, silicon PIN photodiode is used as the detector. This detector could eventually be integrated with the ECL microcell in the same silicon wafer. Our most recent results have demonstrated a detection limit of tens of atto moles of the luminescing label, $\text{Ru}^{+2}(\text{bpy})_3$.

THEORY

The chemicals used in this system are tris- N,N' bipyridyl ruthenium (II) (abbreviated as TBR) and tripropylamine (TPA). The behavior and related properties of solutions containing these two chemicals have been previously reported [1,2]. To summarize, TPA and TBR are electrochemically oxidized at about the same potential, and became TPA^+ and TBR^+ respectively. TPA^+ is highly unstable and becomes quickly deprotonated. The deprotonated TPA^+ (denoted as TPA') is a strong reducing agent, which reduces TBR^+ to TBR. During electron transfer, energy transfer also occurs, resulting in photon emission from the excited TBR. This is the underlying principle for ECL. Another possible reaction is that the TPA' also reduces TBR, producing TBR^- . When TBR^- reacts with TBR^+ , light is also emitted [3]. The wavelength of luminescence from TBR is at 630 nm, i.e. orange. In order to use ECL in the quantification of replicated DNA fragments, TBR is attached to a DNA probe which is hybridized in situ with the DNA of interest during PCR. One method for then "collecting" the TBR labeled DNA is to attach the DNA to magnetic beads, which can be then attracted by an external magnet to the oxidizing electrode. This method, which is currently employed by a commercial instrument (Perkin-Elmer

QPCR 5000), has been tested with our micro-instrument, and the preliminary results are presented here.

MICROINSTRUMENT CONFIGURATION

The microinstrument comprises three silicon substrates, which are vertically assembled to form a small cavity with two, thin film Platinum electrodes, V-groove fill ports, and a detector. A silicon, PIN photodiode (EG&G, S1337-66BR), caps off the cavity, with its sensing surface facing the electrodes. The other silicon layers have been anisotropically etched in KOH, coated with Platinum and bonded together, defining the cell volume (150 microliters) and electrode spacing (500 microns). A cross-section of the microinstrument is shown in figure 1. The bottom Pt electrode is where electrochemical oxidation takes place. The upper Pt electrode has a square opening through which ECL can pass to the detector.

A transimpedance pre-amplifier, chopper circuit and lock-in amplifier are connected to the PIN photodiode as shown in figure 2. An EG&G PARC 273 potentiostat is used to apply a triangle voltage waveform to the Pt electrodes for the electrochemical reaction. The output signal, corresponding to the measured light intensity, is monitored and stored on a digital storage oscilloscope and then transferred to a IBM compatible PC for signal processing.

EXPERIMENTAL DETAILS AND RESULTS

Two experiments were conducted to evaluate the instrument. The first experiment used a set of test solutions containing TBR in concentrations of 1nM, 10nM, 100nM and 1 μ M. Each solution is introduced into the cell with a syringe, followed by measurement, and flushing with water. At a concentration of 1nM, no signal could be distinguished from background noise. However, for 10nM solution, a strong signal can be detected. So the detection limit of this microinstrument configuration for free TBR lies somewhere between 1nM and 10nM. A semilog plot of light intensity vs. concentration is shown in figure 3. The second experiment used TBR labeled magnetic beads, obtained in a test kit from Perkin-Elmer. The kit contains four standards. Standard 1 contains unlabeled magnetic beads, Standard 2 contains magnetic beads bound with 100 attomoles TBR, Standard 3 contains magnetic beads with 1200 attomoles TBR and Standard 4 contains magnetic beads with 12000 attomoles TBR. Each standard has a solution volume of 40ml. All of the magnetic beads in each standard solution were captured and put into 2 ml of phosphate buffer plus TPA. The beads were re-suspended by shaking and then 150 microliters was removed and introduced into the cell. The same measurement procedures were followed as those used with the unbound TBR, except an external magnet was

used to trap the magnetic beads onto the surface of the bottom Pt electrode. In this manner, quantities of TBR less than 100 attomoles, and in the order of 10 attomoles, were detected. The light intensity vs. time plot for bead standards 2 and 3 is shown in figure 4.

ACKNOWLEDGEMENTS

Y.-T. Hsueh acknowledges support from LLNL and a fellowship from the Associated Western Universities.

REFERENCES

1. J. K. Leland and M. J. Powell, J. Electrochem. Soc., 137, pp. 3127-3131 (1990).
2. Y.-T. Hsueh, et. al., 188th Meeting of the Electrochem. Soc., Proceedings of the Symposium on Microstructures and Microfabricated Systems III, pp. 117-122, (1995).
3. I. Rubinstein and A. J. Bard, J. Amer. Chem. Soc., 103, 512-516 (1981).

Figure 1. Cross-section schematic of the three silicon layers which form the electrochemical cell, with detector and external magnet.

Figure 2. Measurement set-up for detection of ECL from $\text{Ru}^{2+}(\text{bpy})_3$.

Figure 3. ECL intensity vs. log conc. of TBR.

Figure 4. Pt voltage and photodiode output vs. time for TBR bound to magnetic beads.