SIMPLE BIOANALYSIS

George M. Whitesides

Department of Chemistry and Chemical Biology Diagnostics for All
Wyss Institute for Biologically Inspired Engineering
Harvard University, Cambridge, MA, USA

ABSTRACT

Diagnosis of disease starts with observation and measurement. Biomedical analyses intended to aid diagnosis are important in determining treatment; they are also important in determining the cost of treatment (for better or worse). The cost of diagnostics is usually hidden from patients (and largely, from doctors), but they can be significant. When diagnostics developed for use in the developed world are transferred into the developing world, where cost is often a key constraint, the issue of cost can be critical in allowing or denying their use to patients. Cost is also important in applications in public health, agriculture and veterinary medicine, and a variety of others.

We are exploring bioanalytical methods that have the characteristics that they provide accurate information that is useful in diagnosis, and do so at costs that are acceptable in the developing world. The core technology, which is being developed in a collaboration between Harvard University and Diagnostics for All (a not-for-profit company focused on engineering development of low-cost diagnostic devices) used paper patterned into hydrophobic and hydrophilic regions, to achieve control of microfluidic flows similar to (or in some cases superior to those generated in) open-channel microfluidic systems. There are also a number of highly developed, low-cost analytical technologies that can be re-purposed for use in low-cost diagnostics: we have focused our developments on glucometers, which provide a convenient method of conducting electroanalyses.

The methods being explored to generate new types of low-cost diagnostics also extend to less familiar types of analyses (for example, two-phase aqueous polymer systems, and magnetic levitation, for density-based separations; 3D mammalian cell growth in gels supported in paper and mesh sheets). The early consideration of “cost” and “simplicity” as parameters in the design of all these systems is, in fact, a very useful stimulus to consider solutions of types not necessarily immediately obvious in an academic environment.

Another key component of this area of diagnostic technology concerns the communication of results to doctors or other capable to carrying out diagnostics, since doctors are also in very limited supply in the developing world. A key initial step in the solution of this problem is the coupling of diagnostic results to (probably, initially, low-capability) cell-phones. We have also carried out initial steps in this area.

Some of this work is described in these papers:

- [1] "Paper-Based ELISA". Cheng, C., Martinez, A. W., Gong, J., Mace, C. R., Phillips, S. T., Carrilho, E., Mirica, K. A. & Whitesides, G. M. *Angew. Chem.* . 2010, 122, 4881-4884
- [2] "Programmable Diagnostic Devices Made from Paper and Tape". Martinez, A. W., Phillips, S. T., Nie, Z., Cheng, C., Carrilho, E., Wiley, B. J. & Whitesides, G. M. *LOC.* 2010, 10, 2499-2504
- [3] "Integration of Paper-Based Microfluidic Devices with Commercial Electrochemical Readers". Nie, Z., Deiss, F., Liu, X., Akbulut, O. & Whitesides, G. M. *LOC.* 2010, 10, 3163-3169 PMC3060706
- [4] "Diagnostics for the Developing World: Microfluidic Paper-Based Analytical Devices". Martinez, A. W., Phillips, S. T., Whitesides, G. M. & Carrilho, E. *Anal. Chem.* 2010, 82, 3 - 10
- [5] "Paper-Based Piezoresistive MEMS Sensors". Liu, X., Mwangi, M., Li, X., O'Brien, M. & Whitesides, G. M. *LOC.* 2011, 11, 2189-2196
- [6] "Measuring Markers of Liver Function Using a Micropatterned Paper Device Designed for Blood from a Fingerstick". Vella, S. J., Beattie, P., Cademartiri, R., Laromaine, A., Martinez, A. W., Phillips, S. T., Mirica, K. A. & Whitesides, G. M. *Anal. Chem.* 2012, 84, 2883-2891
- [7] "Measuring Binding of Protein to Gel-Bound Ligands Using Magnetic Levitation". Shapiro, N. D., Mirica, K. A., Soh, S., Phillips, S. T., Taran, O., Mace, C. R., Shevkoplyas, S. S. & Whitesides, G. M. *J. Amer. Chem. Soc.* 2012, 134, 5637-5646