Consistency of CT KUB Radiation Dose and Exposure Parameters in King Abdulaziz Medical City in Jeddah- Quality Assessment

Rahaf H. Almoutairi ¹, Shahd H. Bugis ¹, Zainab A. Alharbi ¹, Samaher M. Alattas ¹, Essam Banoqitah^{2*}, Khalid M. Alshamrani ¹, Ahmed M. Subahi ⁴, and Shaza S. Alsharif ^{1,3}

¹College of Applied Medical Sciences, King Saud Bin Abdulaziz University for Health Sciences, National Guard, Jeddah, ²Nuclear Engineering Department, Faculty of Engineering, King Abdulaziz University, ³Medical Imaging Department, King Abdulaziz Medical City, National Guard, Jeddah, and ⁴College of Science and Health Professions, King Saud Bin Abdulaziz University for Health Sciences, National Guard, Jeddah, Saudi Arabia

* ebanqitah@kau.edu.sa

Abstract. Computed Tomography (CT) is a type of Medical Imaging that uses X-ray to generate cross-sectional images of the body with an effective dose that is higher than the dose of conventional radiography. For instance, the radiation dose for Kidney, Ureter, and Bladder (KUB) in conventional radiography is 0.7mSv, whereas, in the CT scan is 10mSv. Therefore, this study aimed to evaluate the patient radiation dose delivered in unenhanced CT-KUB examinations using two CT machines in compliance with the American College of Radiology (ACR) standards. This was a retrospective, cross-sectional study of all patients who had unenhanced CT-KUB examinations in King Abdulaziz Medical City in Jeddah between January 1st and June 30th, 2018. A total of 264 patients were included in the study. The mean CTDIvol, DLP, Time (s), pitch and effective dose were significantly different in the CT machines with a p-value between 0.004-<0.001. Both CT machines were within the ACR standards range of 10mSv. There was a positive strong correlation between the BMI (29.6 kg/m²) and the effective dose with p-value=<001. The study further revealed that radiation doses delivered by the two CT scanners are compliant with the ACR standards.

Keywords: Computed Tomography, KUB, Quality, Radiation dose, ACR standards.

1. Introduction

Computed Tomography (CT) has become a mainstay for diagnosing medical diseases. One of the most common examination that can be done with a CT machine is KUB examination, which is the examination for Kidney, Ureter, and Bladder ^[1]. This is because CT has many advantages such as quick, painless, and high-resolution image. However, its main limitation is the amount of radiation delivered to the

patients ^[2]. The concern of this study to calculate and compare the radiation dose that delivered to unenhanced CT KUB patients with American College of Radiology (ACR) standards.

1.1 CT Dose Descriptors

CT dose index (CTDI) is used to estimate the radiation dose outcome from each CT machine ^[2]. Dose index can be classified into CTDI100, CTDIw, and CTDIvol ^[3].

CTDIvol measures the dose in the X-Y-Z axis plane for a single slice. The CTDIvol for helical CT scan is $CTDIvol = CTDIw/Pitch$ [3]. CTDIvol also considered as a standard for comparing and assessing radiation dose among CT machines [2,4]. Dose Length Product (DLP) is also used along with CTDIvol to calculate the radiation dose. DLP represents the total dose regarding the total length of the scan which could be estimated from the number of slices multiplied by the slice width. DLP can be calculated by using the following equation: $DLP = CTDIvol (mGy) * scan length (cm)$ [5]. Moreover, the radiation dose also known as effective dose (EfD) takes into account the sensitivity of different tissues and the absorbed dose to specific organs. The unit of measurement for EfD is the milli-sieverts (mSv). According to the ACR standards, the estimated effective radiation dose for CT KUB examination is 10 mSv or less. The effective dose can be calculated by using this equation $EfD = k * DLP$, while k is the conversion coefficient that is specific only to the anatomical region scanned [6]. There are several factors affect the patient radiation dose, however, in this research, we mainly focused on the pitch, kVp, mAs, slice width and Spacing [2].

2. Methodology

2.1 Studied Population

The study was conducted at the medical imaging department- CT suite in King Abdulaziz Medical City (KAMC) in Jeddah, Saudi Arabia. This retrospective, observational, analytical, cross-sectional study included 264 – 176 male and 88 female - patients who underwent unenhanced KUB examinations, over the six months, from 1st of January 2018 to 30th of June 2018 and involved only adult patients. For all patients, we used these parameters: 100 or 120 kVp, 121-599 mA, pitch 0.984 or 1.374- and 2.5-

mm slice thickness. The Body Mass Index (BMI) range from 19.51 kg/m² to 31.75 kg/m² with average 25.59 kg/m².

2.2 Data Collection Plan and Instrument Used, Measurements

The data obtained from all the patients with renal calculi by utilizing two types of CT machines, which are CT HD, and LightSpeed VCT. CTDIvol and DLP have taken for all unenhanced CT KUB examinations from the hospital information systems (HIS) and pictures archiving and communication system (PACS).

2.3 Evaluation of Radiation Dose

The effective dose is a dose quantity used by the American College of Radiology (ACR) to determine whether an individual's radiation exposure within the accepted range or not, see Table 1 [7]. The calculation of the effective dose was done by utilizing this formula $EfD = DLP * K$ which is adapted from the American Association of Physicists in Medicine (AAPM) [6].

Table 1. ACR Dose Reference for Adult.

Procedure	Approximate Effective radiation dose	Comparable to the natural background radiation for
CT abdomen	10 mSv	3 years

2.4 Data Management and Analysis Plan

IBM SPSS Statistics version 21 was used to analyze the data. Descriptive statistics were conducted. The means and standard deviations for continuous variables were calculated for the whole sample. To compare the mean of each variable by type of CT machine, independent sample T-tests were used. The significant level was set up at P 0.05.

3. Results

A total sample of 264 participated in this study. The sample characteristics were reported in Table 2. The mean age of the participants was 48.1 ± 16.3 , and most of them were male 66.7 %. As showed in Tables 2, 3 and 4, approximately three-quarters of the participants ($n = 199$, 75.5%) were assessed using the GE (light speed VCT), while one-quarter of the sample ($n = 65$, 24.6%) were assessed with GE (CTHD). The mean BMI was 29.59 ± 6.45 .

Table 2. Sample characteristics.

Characteristics	Total sample N= 264	
	Mean	SD
Age (yrs.)	48.1	16.3
CTDI _{vol} (mGy)	14.1	5.99
DLP (mGy*cm)	714.52	349.90
kVp	106.74	9.63
mA	337.20	103.93
Time (s)	753.14	79.07
Pitch	1.13	.17
Slice number	179.67	22.03
Slice thickness (mm)	2.50	.03
length cover (mm)	449.58	55.36
BMI (kg/m ²)	29.59	6.45
Effective dose (mSv)	10.69	5.28

The results of independent sample *T*-tests show significant differences of the mean of six variables based on the type of CT machine, see Table 5. The mean CTDI and DLP significantly differ based on CT machine ($t(262) = 3.77$, $p < .001$; $t(262) = 2.89$, $p < .001$). Besides, mA, Time(s), pitch and effective dose significantly differed when using different CT machine. No significant differences were found in kVp, mAs, slice number, slice thickness and length cover when using a different type of CT machines.

Table 3. Gender distribution.

Gender	Total number of patients	Percentage %
Male	176	66.7%
Female	88	33.3%

Table 4. Patients' distribution.

Type of CT machine	Total number of patients	Percentage %
GE light speed VCT	199	75.5%
GE CT 750HD	65	24.6%

Table 5. Compare the mean of each variable by type of CT machine using independent sample *t* tests.

Variable	Group used GE (light speed VCT) / Main CT (n= 199)		Group used GE (CTHD) (n= 65)		Mean difference between groups	t	df	P value
	Mean,	SD	Mean,	SD				
Age (yrs.)	48.1	16.3	48.28	16.68	.804	.345	262	.73
CTDI _{vol} (mGy)	14.1	5.99	14.86	6.04	3.15	3.77	262	<.001
DLP (mGy*cm)	714.52	349.90	749.69	358.65	142.82	2.89	262	.004
kVp	106.74	9.63	106.63	9.65	-.44	-.32	262	.75
mA	337.20	103.93	330.07	106.36	-28.98	-2.09	122.41	.04
Time(s)	753.14	79.07	773.82	77.38	84.01	8.35	262	<.001
Pitch	1.13	.17	.99	.086	-.2572399	-14.32	262	<.001
Slice number	179.67	22.03	178.99	21.36	-2.7843	-.88	262	.38
Slice thickness (mm)	2.50	.03	2.50	.04	.01	.67	262	.50
length cover (mm)	449.58	55.36	448	53.79	-6.42	-.81	262	.42
BMI (kg/m ²)	29.59	6.45	29.74	6.84	.63	.68	262	.49
Effective dose (mSv)	11.21	5.42	9.10	4.48	2.11	2.84	262	.005

There was a positive strong correlation between the BMI and the effective dose using Pearson correlation coefficients, $r = 0.502$, $n = 264$, $p < .001$ (see Fig. 1).

Correlations		BMI	effective dose
BMI	Pearson Correlation	1	.502**
	Sig. (2-tailed)		.000
	N	264	264
effective dose	Pearson Correlation	.502**	1
	Sig. (2-tailed)	.000	
	N	264	264

Correlation is significant at the 0.01 level (2-tailed).

Fig. 1. Pearson correlation coefficients between BMI and effective dose.

The analyzing data shows that there was an effect of different values of pitch on the patient radiation dose. For example, when the pitch was at 0.984, the patient radiation dose was more in the light speed scanner (See Fig. 2 and 3).

4. Discussion

Our results showed significant differences in the mean of six variables based on the type of CT machine. The mean of CT dose index (CTDI), DLP, mA, Time(s), pitch, and effective dose significantly differ as shown in Table 2. According to the American College of Radiology (ACR), the effective dose of abdomen region should not exceed 10

mSv, and this value is approximately close to the total mean of two CT machines which is 10.69. However, this small difference between the two values is accepted due to the variation in patients' sizes and BMIs. The recommended value usually is applicable for a patient with average weight and BMI. Patients in our study vary in their sizes, so we need to adjust the dose parameters for each patient. Estimating Effective Dose for CT Using DLP Compared with Organ Doses is a previous study that compared the mean of the effective dose of the abdomen region based on the International Commission on Radiological Protection ICRP 26, 60, 103, instead of ACR. The findings were close to the ICRP standards with slight differences that might also be related to the variation in patients' size [8].

Fig. 2-3. The Effect of different values of pitch on the patient radiation dose.

4.1 Limitations

There are a few limitations that affect our findings. First, the number of patients is not equal between the two CT machines. Second, the exclusion of the comparison between the phantoms of each CT scanner minimizes the accuracy of our results.

4.2 Strength

This type of study has been done at KAMC-JD for the first time.

5. Conclusion

CTHD scanner gives less radiation than the LightSpeed VCT scanner. The radiation doses delivered by the two CT machines in KAMC-JD are compliant with the ACR standards. For further improvements, we recommend increasing the accuracy of the study by increasing the patient's sample size, inclusion the comparison between the phantoms, and calculating the effective dose of each organ instead of the abdomen region. Also, apply this research study on different CT procedures and not only unenhanced CT KUB examination. Moreover, expand the comparison to include two or more hospitals instead of one.

References

- [1] **Nadeem, M., Ather, M., Jamshaid, A., Zaigham, S., Mirza, R. and Salam, B.**, Rationale use of unenhanced multi-detector CT (CT KUB) in evaluation of suspected renal colic. *International Journal of Surgery* [Internet]. 2012 [cited 16 March 2018]; 10(10):634-637. Available from: [http://www.journal-surgery.net/article/S1743-9191\(12\)00817-5/pdf](http://www.journal-surgery.net/article/S1743-9191(12)00817-5/pdf)
- [2] **Romans, L.**, Computed tomography for technologists. 1st ed. Philadelphia: Wollters Kluwer Health; 2011
- [3] **McCollough, C., Leng, S., Yu, L., Cody, D., Boone, J., and McNitt-Gray, M.**, CT Dose Index and Patient Dose: They Are Not the Same Thing. *Radiology* [Internet]. 2011 [cited 18 March 2018]; 259(2):311-316. Available from: <http://pubs.rsna.org/doi/pdf/10.1148/radiol.11101800>
- [4] **Sung, M.**, Current status of low dose multi-detector CT in the urinary tract. *World Journal of Radiology* [Internet]. 2011 [cited 17 March 2018]; 3(11):256. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC3226959/>
- [5] **McNitt-Gray, M.**, AAPM/RSNA Physics Tutorial for Residents: Topics in CT. *RadioGraphics* [Internet]. 2002 [cited 17 March 2018]; 22(6):1541-1553. Available from: <https://pubs.rsna.org/doi/pdf/10.1148/rg.226025128>
- [6] **Aapm.org.** (2018). [online] Available at: <https://www.aapm.org/meetings/amos2/pdf/34-9723-93678-499.pdf> [Accessed 23 Dec. 2018].
- [7] **Acr.org.** (2018). [online] Available at: <https://www.acr.org/-/media/ACR/Files/Radiology-Safety/Radiation-Safety/Dose-Reference-Card.pdf>
- [8] **Home (AJR)** [Internet]. *Ajronline.org*. 2018 [cited 7 December 2018]. Available from: <https://www.ajronline.org/doi/pdf/10.2214/AJR.09.3462>

تناسق الجرعة الإشعاعية للتصوير المقطعي للكلية والحالب والمثانة في مدينة الملك عبدالعزیز الطبية بجدة - تقييم الجودة

رهف حمید المطيري^١، و شهد حسن بوقس^١، و لما الخوتاني^١، و زینب علی الحربي^١، و عصام بانقیطة^٢،
و خالد مسفر الشمراني^١، و أحمد سبجي^٣، و شذى سالم الشریف^{٣،١}

^١ كلية العلوم الطبية التطبيقية، جامعة الملك سعود بن عبد العزيز للعلوم الصحية، الحرس الوطني، جدة، و ^٢ قسم الهندسة النووية،
كلية الهندسة، جامعة الملك عبد العزيز، ص. ب. ٨٠٢٠٤، جدة ٢١٥٨٩، و ^٣ قسم التصوير الطبي، مدينة الملك عبد العزيز
الطبية، الحرس الوطني، جدة، و ^٤ كلية العلوم والمهن الصحية، جامعة الملك سعود بن عبد العزيز للعلوم الصحية، الحرس
الوطني، جدة، المملكة العربية السعودية

ebanoqitah@kau.edu.sa

المستخلص. تُعد الأشعة المقطعية نوعًا من أنواع التصوير الطبي. وتظهر ميزة التصوير
المقطعي بجمع سلسلة من صور الأشعة السينية المأخوذة من زوايا مختلفة حول الجسم. لتظهر
الجسم بشكل ثلاثي الأبعاد. كما للتصوير المقطعي مزايا، فإن له عيوبًا أيضًا، فالتصوير
المقطعي يستعمل كمية كبيرة من الإشعاع مقارنة بالأشعة السينية. فعلى سبيل المثال، الجرعة
الإشعاعية عند تصوير الكلية والحالب والمثانة باستعمال الأشعة السنية تصل إلى ٠,٧ ملي
سيفرت مقارنة بالأشعة المقطعية فالجرعة تصل إلى ١٠ ملي سيفرت. وتهدف هذه الدراسة
لحساب الجرعة الإشعاعية للمرضى بمدينة الملك عبدالعزيز الطبية الخاضعين لفحص الكلية
والحالب والمثانة لمقارنة الجرعة الإشعاعية لهم بالمقياس المسموح من الجمعية الأمريكية للأشعة.
وقد تم عمل دراسة مستعرضة استعادية لجميع المرضى الخاضعين لفحص الأشعة المقطعية
للكلية والحالب والمثانة بمدينة الملك عبدالعزيز الطبية خلال الفترة من الأول من شهر يناير
وحتى يوم ثلاثين من شهر يونيو. وتضمنت الدراسة ٢٦٤ مريضًا، وكان المتوسط الحسابي لكل
من CTDIvol و DLP و الوقت و Pitch والجرعة الفعالة مختلف بشكل كبير بين أجهزة تصوير
الأشعة المقطعية ذات القيمة الاحتمالية بين ٠,٠٠٤ - ٠,٠٠١٦. ومن خلال الدراسة تم
التوصل إلى أن كل من جهازي التصوير المقطعي كانا ضمن نطاق المعايير التابعة للجامعة
الأمريكية للأشعة، والذي يصل إلى حوالي ١٠ ملي سيفرت. ونستخلص من هذا أن جرعات
الأشعة التي قدمها جهازي تصوير الأشعة المقطعية كانت متوافقة مع الجرعة الإشعاعية
المسموح بها من الجمعية الأمريكية للأشعة.

كلمات مفتاحية: التصوير المقطعي، KUB، الجودة، جرعة الإشعاع، معايير الجمعية الأمريكية
للأشعة.