Supplement of

ACTRIS ACSM intercomparison – Part 2: Intercomparison of ME-2 organic source apportionment results from 15 individual, co-located aerosol mass spectrometers

R. Fröhlich et al.

Correspondence to: A. S. H. Prévôt (andre.prevot@psi.ch)

z -score:

$$z = \frac{(x_{i,j} - X_j)}{\sigma_j} \quad (\text{Eq. S1})$$

with $x_{i,j}$: contribution of source j for instrument i

X_j : median contribution of source j

5 σ_j : standard deviation

Tab. S 1. Participating ACSM instruments sorted by serial number. The order of instruments in this table purposely does not coincide with the order #1 - #13 used in the text to avoid an unintentional “rating” of individual instruments.

ACSM serial #	Operating institution(s)	Country	Monitoring location during ACTRIS
A140-104	University of Helsinki	Finland	Hyytiälä, FIN
A140-110	IDAIA-CSIC	Spain	Montseny, ESP
A140-113	LSCE	France	Gif-Sur-Yvette (SIRTA), FRA
A140-133	National University of Ireland, Galway	Ireland	Mace Head, IRL
A140-134	DWD	Germany	Hohenpeißenberg, DEU
A140-142	LSCE	France	Cap Corse, FRA
A140-143	TROPOS	Germany	Melpitz, DEU
A140-144	NILU	Norway	Birkenes, NOR
A140-145	PSI	Switzerland	Cabauw, NLD
A140-151	EC-JRC & ENEA	Italy & Italy	Ispra, ITA
A140-152	ISAC-CNR	Italy	Bologna, ITA
A140-153	CIEMAT	Spain	Madrid, ESP
A140-156	King’s College London	United Kingdom	North Kensington, GBR
ToF A003	PSI	Switzerland	Jungfrauoch, CHE
HTOF AMS 215-081	Ecole Nationale Supérieure des Mines de Douai	France	Revin, FRA (EMEP)

Tab. S 2. Calibrated values of the Q-ACSM response function (ions/pg value for the ToF-ACSM) and applied fitting method for the relative ion transmission (RIT) curve of the quadrupole mass spectrometer.

ACSM serial #	Response factor (RF) in A/($\mu\text{g}/\text{m}^3$)
A140-104	3.0×10^{-11}
A140-110	5.3×10^{-11}
A140-113	4.4×10^{-11}
A140-133	4.0×10^{-11}
A140-134	3.1×10^{-11}
A140-142	3.7×10^{-11}
A140-143	5.5×10^{-11}
A140-144	2.8×10^{-11}
A140-145	2.4×10^{-11}
A140-151	3.0×10^{-11}
A140-152	2.7×10^{-11}
A140-153	3.1×10^{-11}
A140-156	3.3×10^{-11}
ToF A003	8.6 ions/pg

Tab. S 3. Modifications of calculated ME2 input matrices before analysis. m/z 12 was removed in all Q-ACSM instruments due to non-physical negative signals.

Instrument	m/z downweighted by factor 3	additional info
ToF	none	
#1	m/z 13 & 24	
#2	m/z 13 & 15	
#3	m/z 13	
#4	m/z 13, 15, 24 & 37	
#5	none	
#6	m/z 15	m/z 13 removed (neg. S/N)
#7	m/z 13	
#8	m/z 13, 15, & 37	
#9	m/z 13 & 15	
#10	m/z 13 & 15	
#11	none	
#12	none	
#13	m/z 13 & 15	
HR-AMS	CH_2^+ (14.0156) & C_3H_3^+ (39.0235) by factor 3 CH_3O_3^+ (62.9905) by factor 6	7 entries removed: CH_4^+ (16.0313), CH_4O^+ (32.0262), CH_2O_2^+ (46.0055), CH_4NO^+ (46.0293), $\text{C}_2\text{H}_6\text{O}^+$ (46.0419), C_2HCl_2^+ (94.9455), $\text{C}_3\text{H}_3\text{O}_5^+$ (118.998)

Tab. S 4. Correlations (R^2) between factors from HR PMF analysis with external measurements.

R^2	BC_{wb}	BC_{ff}	NO_x	SO_4
#1 (HOA)	0.39	0.68	0.65	0.01
#2 (COA)	0.64	0.40	0.29	0.12
#3 (OOA)	0.20	0.15	0.06	0.43
#4 (BBOA)	0.90	0.27	0.22	0.08

Fig. S 1. (a) Median organic mass spectrum of the 13 Q-ACSMs. The signals at $m/z = 44$ and all related sticks ($m/z = 16, 17$ and 18) were removed and the mass spectrum was then renormalised. The boxes represent the interquartile range (IQR) (25-75 percentile) for each m/z stick and the whiskers represent the corresponding full range over all instruments. The colour bar represents the ratio of the width of the individual boxes in relation to the corresponding median in percent. (b) Fractions of the total organic signal at single m/z channels for all 15 participating instruments after all CO_2 related peaks were removed from the spectrum. The instruments were sorted like in Fig. 2a (by fraction of $m/z = 44$ before removing). Grey: f29, blue: f43, purple: f55, pink: f57, red: f60. The respective fractions are given as numbers in the same colours.

Fig. S 2. Ratio of f44 between a Q-ACSM with high f44 (#12) and a Q-ACSM with low f44 (#2, cf Fig. 2b) plotted against concentration (black: total, yellow: ammonium, green: organics, blue: nitrate, red: sulphate). Slopes of an orthogonal linear regression are given in each plot. Slopes are all positive but very small, indicating only a very minor dependency of f44 on concentration.

Fig. S 3. Fractions of m/z 43 and 44 in the organic spectra of three different filter samples (z1, z22 and z49). All three filters contain ambient aerosol particles sampled in Zurich, CH. Samples z1 and z22 were taken in winter and sample z49 was taken in summer. The filters were extracted into a liquid solution and subsequently nebulised to be able to measure them with an ACSM. For method description refer to Dällenbach et al. (2014). (a) f44, f43 and ratio of f44 vs f43 at different vaporiser temperature settings for the three different filters. Each spectrum was averaged for 3×10 min. (b) Average f44, f43 and ratio of f44 vs f43 over the measurements with vaporiser temperatures between 550 °C and 660 °C (possible error margin of the ACSM temperature reading). The error bars show the first standard deviation.

Tab. S 5. Correlations (R^2) between factors from three factor PMF analysis with external measurements.

R^2	HOA - BC _{ff}	HOA - NO _x	BBOA - BC _{wb}	OOA - SO ₄
ToF	0.41	0.49	0.88	0.51
AMS (UMR)	0.56	0.60	0.86	0.56
#1	0.67	0.62	0.93	0.72
#2	0.58	0.60	0.93	0.52
#3	0.44	0.44	0.91	0.71
#4	0.63	0.66	0.93	0.54
#5	0.53	0.60	0.86	0.76
#6	0.53	0.52	0.87	0.73
#7	0.43	0.46	0.87	0.72
#8	0.61	0.54	0.88	0.72
#9	0.49	0.48	0.85	0.75
#10	0.62	0.70	0.91	0.75
#11	0.64	0.61	0.84	0.74
#12	0.52	0.54	0.91	0.78
#13	0.53	0.53	0.86	0.59

Fig. S 4. Time series correlations between factors from HR PMF analysis with external measurements during the whole two weeks of experiment. Corresponding R^2 values are given in the plot. (a) HOA factor (grey) plotted together with NO_x (black) measured with the NO_x analyser and BC_{ff} (red) time trace from the BC source apportionment of the aethalometer data. (b) COA-like factor (yellow) plotted together with time trace of $\text{C}_3\text{H}_3\text{O}^+$ ion (pink) and $\text{C}_6\text{H}_{10}\text{O}^+$ ion (grey) from the AMS HR analysis. Coefficients of determination (R^2) with COA-like and HOA factors are given in the plot. (c) BBOA factor (brown) plotted together with time traces of gaseous methanol (violet) and acetonitrile (red), both measured by the PTR-MS and BC_{wb} (black) from the BC source apportionment of the aethalometer data. (d) OOA factor (green) plotted together with inorganic nitrate (blue) and sulphate (red) measured by the HR-AMS.

Fig. S 5. High-resolution COA factor spectrum from Crippa et al. (2013) (top) and from this study (bottom).

Fig. S 6. Relative contributions of HOA (grey), OOA (green) and BBOA (brown) factors from a three factor PMF solution without constraints for all 15 instruments (sorted by total f44). The shown AMS solution is from the analysis of the UMR AMS data set. Average contributions and the first standard deviation are given in the legend.

Fig. S 7. Full absolute time series of all four ME2 factors in $\mu\text{g}/\text{m}^3$ (from top to bottom: HOA, COA-like, BBOA and OOA) normalised to the same total organic concentrations (reference: HR-ToF-AMS). Pink: HR-ToF-AMS, green: ToF-ACSM, grey: all 13 Q-ACSMs.

Fig. S 8. Correlations of normalised HOA factor time series (reference: total organic mass of HR-ToF-AMS) of all instruments (name indicated in the y label) to the median of all instruments. Slopes and coefficients of correlation (R^2) are given in the plots. The black dashed line indicates the 1 : 1 line.

Fig. S 9. Correlations of normalised COA factor time series (reference: total organic mass of HR-ToF-AMS) of all instruments (name indicated in the y label) to the median of all instruments. Slopes and coefficients of correlation (R^2) are given in the plots. The black dashed line indicates the 1 : 1 line.

Fig. S 10. Correlations of normalised BBOA factor time series (reference: total organic mass of HR-ToF-AMS) of all instruments (name indicated in the y label) to the median of all instruments. Slopes and coefficients of correlation (R^2) are given in the plots. The black dashed line indicates the 1 : 1 line.

Fig. S 11. Correlations of normalised OOA factor time series (reference: total organic mass of HR-ToF-AMS) of all instruments (name indicated in the y label) to the median of all instruments. Slopes and coefficients of correlation (R^2) are given in the plots. The black dashed line indicates the 1:1 line.

Fig. S 12. ME2 source spectra of the ToF-ACSM solution (sticks). The boxes and whiskers show the IQR and the full range of the Q-ACSM spectra, respectively. The line in the box indicates the median of all Q-ACSM. From top to bottom: HOA, COA-like, BBOA and OOA.

Tab. S 6. Average factor concentrations with first standard deviation in $\mu\text{g}/\text{m}^3$. This combines the uncertainties of the source apportionment with the uncertainties of the total organic mass discussed in part 1 of this study by Crenn et al. (2014).

factor	average concentration ($\mu\text{g}/\text{m}^3$)	first standard deviation ($\mu\text{g}/\text{m}^3$)
HOA	1.01	0.25
COA-like	1.06	0.35
OOA	2.90	0.51
BBOA	2.11	0.73

Fig. S 13. Source profiles from ME-2 "best solutions" of all 13 Q-ACSMs. All normalised to 1. COA-like (yellow) and HOA (grey). Organised from left to right and top to bottom according to the instrument numbers #1 - #13.

Fig. S 14. Source profiles from ME-2 "best solutions" of all 13 Q-ACSMs. All normalised to 1. OOA (green) and BBOA (brown). Organised from left to right and top to bottom according to the instrument numbers #1 - #13.

Fig. S 15. z -score values for all factors and all instruments. z -score values were calculated with median contribution as reference and standard deviations determined with the robust algorithm A of ISO 13528 according to Eq. S1. $|z| \leq 1$: "ok", $1 < |z| \leq 2$: "acceptable", $2 < |z| \leq 3$: "warning", $|z| > 3$: "action" (see §7.4.2, ISO13528 (2005)).

Update of the Q-ACSM PMF export function in recent analysis software update (version 1.5.5.0) and its influence on the presented dataset

After the presented work was performed an update of the analysis software was released introducing changes of the PMF matrix calculation procedure. While in the former version the same relative ion transmission (RIT) curve (red curve in Fig. S16a) was applied to all datasets, the updated version introduces an individual correction (blue curve in Fig S16a) for each instrument with the goal to reduce spectral differences between instruments. All data discussed in this manuscript, however, was calculated with version 1.5.3.2, i.e. with the same RIT applied. Fig. S16a shows both RITs of the instrument (#9) with the highest deviation between “old” and “new” calculation of all 13 Q-ACSMs. Note the error range for the calculated curve. In almost all cases (except #10 & #13) the “new” calculation leads to lower intensities of the m/z channels >55 . The error matrix calculation is equally affected, however there the change is about a factor two lower. This means the S/N of the $m/z >55$ was slightly overestimated in the data presented in the manuscript compared to if the data was calculated with the “new” procedures. Figs. S16b shows relative changes of PMF organic matrix and error matrix and Fig. S16c shows the absolute change of the spectrum for the most severely affected instrument #9.

What are possible consequences on the presented work? On the one hand the presented uncertainties may be slightly lower if the “new” calculation was applied because input spectra are more similar, but on the other hand a lower S/N at higher m/z may lead to slightly larger residuals and hence to slightly larger deviations between solutions. A comparison of source apportionment results of one instrument with matrices calculated with both software versions showed no significant differences in either factor contributions, factor time series nor factor profiles.

All Q-ACSM source apportionment results published so far used the “old” calculation method.

Fig. S 16. (a) Relative ion transmission (RIT) curves of instrument #9 with the largest deviation between default and calculated correction (red and blue curve). Black dots and error bars were calculated from naphthalene fragments in the background spectra. Red curve: default correction used in analysis version 1.5.3.2, i.e. for all data presented in this manuscript. Blue curve: calculated correction from fit to the measured data. This correction is applied in version 1.5.5.0 of the analysis software. (b) Relative change of PMF organic matrix (red) and PMF error matrix (black) as function of m/z for the two software versions. Note that the relative change of the error is smaller than the relative change of the signal, leading to decreased S/N in this case. (c) Spectrum (one datapoint) of instrument #9 calculated with the two versions. Red was calculated with version 1.5.5.0, black with 1.5.3.2. Intensities of from the newer software version are slightly lower.

References

- 45 Crenn, V., Sciare, J., Croteau, P. L., Favez, O., Verlhac, S., Belis, C. A., Fröhlich, R., Aas, W., Äijälä, M., Alastuey, A., Artiñano, B., Baisnée, D., Baltensperger, U., Bonnaire, N., Bressi, M., Canagaratna, M., Canonaco, F., Carbone, C., Cavalli, F., Coz, E., Cubison, M. J., Gietl, J. K., Green, D. C., Heikkinen, L., Lunder, C., Minguillón, M. C., Močnik, G., O'Dowd, C. D., Ovadnevaite, J., Petit, J.-E., Petralia, E., Poulain, L., Prévôt, A. S. H., Priestman, M., Riffault, V., Ripoll, A., Sarda-Estève, R., Slowik, J., Setyan, A., and Jayne, J. T.: ACTRIS ACSM Intercomparison: part I - Intercomparison of concentration and fragment results from 13 individual co-located aerosol chemical speciation monitors (ACSM), submitted to *Atmos. Meas. Tech. Disc.*, 2014.
- 50 Crippa, M., DeCarlo, P. F., Slowik, J. G., Mohr, C., Heringa, M. F., Chirico, R., Poulain, L., Freutel, F., Sciare, J., Cozic, J., Di Marco, C. F., Elsasser, M., Nicolas, J. B., Marchand, N., Abidi, E., Wiedensohler, A., Drewnick, F., Schneider, J., Borrmann, S., Nemitz, E., Zimmermann, R., Jaffrezo, J.-L., Prévôt, A. S. H., and Baltensperger, U.: Wintertime aerosol chemical composition and source apportionment of the organic fraction in the metropolitan area of Paris, *Atmos. Chem. Phys.*, 13, 961–981, 2013.
- 65 Dällenbach, K. R., Bozzetti, C., Křepelová, A., Canonaco, F., Wolf, R., Huang, R.-J., Zotter, P., Crippa, M., Slowik, J. G., Zhang, Y., Szidat, S., Baltensperger, U., Prévôt, A. S. H., and El Haddad, I.: Characterization and source apportionment of organic aerosol using offline aerosol mass spectrometry, in prep. *Atmos. Meas. Tech.*, 2014.
- 70 ISO13528: Statistical Methods for Use in Proficiency Testing by Interlaboratory Comparisons, ISO 13528, International Organization for Standardization, Geneva, Switzerland, 2005.